UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/723,720	11/26/2003	Nicholas Popadiuk	ETH-5113	9774
27777 7590 02/26/2007 PHILIP S. JOHNSON JOHNSON & JOHNSON			EXAMINER	
			POUS, NATALIE R	
	N & JOHNSON PLAZ WICK, NJ 08933-7003	A	ART UNIT	PAPER NUMBER
	·		3731	
SHORTENED STATUTOR	Y PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE	
31 DAYS		02/26/2007	PAPER	

## Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

	Application No.	Applicant(s)			
	10/723,720	POPADIUK ET AL.			
Office Action Summary	Examiner	Art Unit			
	Natalie Pous	3731			
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA  - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication.  - If NO period for reply is specified above, the maximum statutory period w  - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tim vill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status					
<ul> <li>1) Responsive to communication(s) filed on 30 De</li> <li>2a) This action is FINAL 2b) This</li> <li>3) Since this application is in condition for allowant closed in accordance with the practice under E</li> </ul>	action is non-final. nce except for formal matters, pro	•			
Disposition of Claims					
4) ⊠ Claim(s) 1-15 is/are pending in the application.  4a) Of the above claim(s) is/are withdraw  5) □ Claim(s) is/are allowed.  6) □ Claim(s) is/are rejected.  7) □ Claim(s) is/are objected to.  8) ⊠ Claim(s) 1-15 are subject to restriction and/or expressions.	vn from consideration.				
Application Papers					
9) The specification is objected to by the Examine	r.				
10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.					
Applicant may not request that any objection to the	drawing(s) be held in abeyance. See	∍ 37 CFR 1.85(a).			
Replacement drawing sheet(s) including the correcting 11) The oath or declaration is objected to by the Expression 11.		, ,			
Priority under 35 U.S.C. § 119					
12) Acknowledgment is made of a claim for foreign a) All b) Some * c) None of:  1. Certified copies of the priority documents 2. Certified copies of the priority documents 3. Copies of the certified copies of the prior application from the International Bureau * See the attached detailed Office action for a list of	s have been received. s have been received in Application ity documents have been received i (PCT Rule 17.2(a)).	on No ed in this National Stage			
	•	•			
Attachment(s)					
Notice of References Cited (PTO-892)     Notice of Draftsperson's Patent Drawing Review (PTO-948)	4) Interview Summary Paper No(s)/Mail Da				
Notice of Draitsperson's Patent Drawing Review (P10-948)     Information Disclosure Statement(s) (PTO/SB/08)     Paper No(s)/Mail Date	5) Notice of Informal P 6) Other:				

## **DETAILED ACTION**

## Election/Restrictions

Examiner acknowledges election of group 1, drawn to apparatus and the withdrawal of claims 16-24, however upon further consideration, examiner has concluded that this application and group 1 contains claims directed to the following patentably distinct species:

Species 1: a prosthetic repair device comprising layers of either non-absorbable material or absorbable materials

Species 2: a prosthetic repair device comprising a nonabsorbable porous material encapsulated with absorbable components

The species are independent or distinct because they require different technical features. For instance, they require separate structural configurations, species 1 being layers, and species 2 being a porous material encapsulated by another material.

Applicant is required under 35 U.S.C. 121 to elect a single disclosed species for prosecution on the merits to which the claims shall be restricted if no generic claim is finally held to be allowable. Currently, no claims generic.

Applicant is advised that a reply to this requirement must include an identification of the species that is elected consonant with this requirement, and a listing of all claims readable thereon, including any claims subsequently added. An argument that a claim is allowable or that all claims are generic is considered nonresponsive unless accompanied by an election.

Application/Control Number: 10/723,720

Art Unit: 3731

Upon the allowance of a generic claim, applicant will be entitled to consideration of claims to additional species which depend from or otherwise require all the limitations of an allowable generic claim as provided by 37 CFR 1.141. If claims are added after the election, applicant must indicate which are readable upon the elected species.

MPEP § 809.02(a).

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.

Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim

Application/Control Number: 10/723,720 Page 4

Art Unit: 3731

remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Natalie Pous whose telephone number is (571) 272-6140. The examiner can normally be reached on Monday-Friday 8:00am-5:30pm, off every 2nd Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhtuan Nguyen can be reached on (571) 272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

NRP 2/15/07

ANHTUAN T. NGUYEN
SUPERVISORY PATENT EXAMINER